

Foam Tipped Applicator

Polypropylene shaft



- Soft foam tip

- Polypropylene shaft

Actual Size

Description

This foam tipped applicator is produced with reticulated 100ppi Medical Grade foam. It is bonded securely to a blue polypropylene shaft with thermo-plastic adhesive.

Dimensions

6" Overall Length (152.4mm)

Tip

- .594" Length (15.09mm)
- .188" Width (4.78mm)

Shaft

- 5.922" Length (150.42mm)
- .120" Dia. (3.05mm)

Qualities

- Lint and residue free
- Soft, non-abrasive tip
- Can be used with IPA and DI water, or dry
- Flexible paddle tip

Standards

- ISO 9001:2008
- FDA GMP's

Applications

- Applying adhesives or other solutions
- Removing dirt, contamination, fluxes or resins
- General cleaning

Ref #	Packaging
1876-PF	20/50 (20 resealable poly bags per case, 50 swabs per bag)
1876-PFB	2/500 bulk (2 poly bags per case, 500 swabs per bag)



Quality Medical Products Since 1919

Puritan Medical Products Company LLC
Guilford, Maine 04443-0149 USA

Tel 800-321-2313 (US and Canada Only)
207-876-3311

Fax 800-323-4153
207-876-3130

Web www.puritanmedproducts.com

Email sales@puritanmedproducts.com

1-800-321-2313
puritanmedproducts.com

Foam Tipped Applicator

Polypropylene Shaft

1876-PF / 1876-PFB

NVR

NON-VOLATILE RESIDUE

Test method IDEMA method M7-98

NVR Solvent - Hexane

NVR 488 µg/swab

The minimum detection limit for NVR is 10 µg/swab

ION CHROMATOGRAPHY TESTING RESULTS

IONIC CLEANLINESS

Test method IDEMA M13-99 (anions) M12-99 (cations)

ANIONS

Fluoride	ND	µg/swab
Chloride	0.6320	µg/swab
Phosphate	ND	µg/swab

CATIONS

Lithium	ND	µg/swab
Sodium	0.1080	µg/swab
Ammonium	0.0280	µg/swab
Potassium	0.0600	µg/swab
Magnesium	ND	µg/swab
Calcium	0.0960	µg/swab

ND - Non Detected

Test procedures available upon request.



Quality Medical Products Since 1919

Puritan Medical Products Company LLC
Guilford, Maine 04443-0149 USATel 800-321-2313 (US and Canada Only)
207-876-3311Fax 800-323-4153
207-876-3130Web www.puritanmedproducts.comEmail sales@puritanmedproducts.com1-800-321-2313
puritanmedproducts.com