

## Specification Sheet



STERILE PAIRS

Nitrile Examination Gloves

Powder-Free

## **Extended Cuff • Ambidextrous**



NitriDerm is manufactured in compliance with the following international standards

Designation	STANDARD		
ASTM D6319-00a(2005)	Standard Specification for Nitrile Examination Gloves for Medical Application		
ASTM D5151-06	Standard Test Method for Detection of Holes in Medical Gloves		
ASTM F1671-03	Standard Test Method for Resistance of Materials Used in Protective Clothing to Penetration by Blood-Borne Pathogens		

Size	ltem Number	Total Length +/- 5mm	Thickness Finger mil	Thickness Finger mm
Small	112100	300	5.1	0.13
Medium	112200	300	5.1	0.13
Large	112300	300	5.1	0.13
X-Large	112350	300	5.1	0.13

Tensile Strength & Elongation	Before Aging	After Accelerated Aging
Tensile Strength Min. (Mpa)	14	14
ASTM Requirement Min. (Mpa)	14	14
Elongation Minimum Min. (%)	500	400
ASTM Requirement Min. (%)	500	400







Innovative Healthcare Corporation is certified to both ISO 9001:2000 quality management system and ISO 13485:2003 quality management system for medical devices.

