ш

trotter padded headrest wing

Item # TR 8022



EU Authorized Representative Drive Medical LTD Ainley's Industrial Estate Elland, West Yorkshire, United Kingdom HX5 9JP

Wenzellte Re/hab

www.drivemedical.com

Wenzellte Re/hab

www.drivemedical.com

installation

- Open Fastex[™] clip (A).
- Place headrest wing on Trotter back (or on headrest extension) and wrap straps around the back.
- 3. Close clip and adjust strap length to fit snugly.



one year limited warranty

Your Drive brand product is warranted to be free of defects in materials and workmanship for one year from the date of the original consumer purchase.

This device was built to exacting standards and carefully inspected prior to shipment. This One Year Limited Warranty is an expression of our confidence in the materials and workmanship of our products and our assurance to the consumer of dependable service.

In the event of a defect covered by this warranty, we will, at our option, repair or replace the device.

This warranty does not cover device failure due to owner misuse or negligence, or normal wear and tear. The warranty does not extend to non-durable components such as rubber accessories, casters, and grips, which are subject to normal wear and need periodic replacement.

If you have a question about your Drive device or this warranty, please contact an authorized Drive dealer.



Wenzel ite Re/hab

© 2006 Medical Depot, Inc. All rights reserved.

Drive is a trademark of Medical Depot, Inc.

Port Washington N.Y. 11050 USA Made in China